



ADVATEK SYSTEMS INC.
C.P./ P.O. Box 966, Moncton, N.B., E1C 8N8
(506) 857-0909 Fax (506) 858-9271

Work Order #

Nº 10145

Customer P.O.#

26727

Certificate #

| | |
|------------------------------------|---------------------------|
| Company Ganong Brothers Chocolate | Date Started Dec. 8, 2007 |
| Address 1 Chocolate Drive | Date Completed |
| City Saint Stephen, NB Postal Code | Make Toledo |
| Contact Fred Mann Ship Via | Model Panther Panel Mount |
| Telephone 506-467-1983 Fax | S/N 5065620-5DA |

| | |
|---|-------------------------------|
| Symptoms: Unit flooded with Sugar and water | Capacity: 1500 kg X .5 |
| when pump malfunctioned, damaged unit. | Approval: Not Legal for Trade |
| Repairs Performed: Check unit, main PCB defective, cannot repair, order new PCB plus 4-20ma PCB and two spare load cells. | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |

Parts Installed:

| Qty | | Unit Cost | Amount |
|---|-----------------------------|------------------|--------|
| | | | |
| | | | |
| | | | |
| | 562 kms. return trip @ 0.90 | 505.80 | 505.80 |
| | Toll Bridge | .50 | .50 |
| | Meal | 7.38 | 7.38 |
| Call In Saturday | | Total Parts | |
| Labour : 2.5 Hours Travel: 6 Hours Total: 8.5 Hours Rate: | | Total Hours | |
| | | Sub | |
| | | HST / GST | |
| Technician: Phil Beatty | | Contract Review: | TOTAL |