

DEPARTMENT OF TRADE AND COMMERCE STANDARDS DIVISION

OTTAWA May 31st, 1951.

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein have been listed as Class A devices, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Equal arm scales with over and under end tower. Models 1301, 1302, 1304, 1305 and 1302R, manufactured by the Exact Weight Scale Company, Columbus, Ohio.

Rating of Apparatus: 150 lbs. capacity.

Application: Weighing of commodities in general trade.

Conditions: As prescribed in P.C. 3459. The over and under chart may be graduated but weight values must not be marked thereon.

Description: The models listed are of the same fundamental design except that the platters vary in size and shape. All models may be fitted with rubber wheels for portability or the casing may be provided with a flange so that the scale casing may be lowered into the floor or work bench. The lever, casing and lower portion of the tower is cast iron. Pivots and bearings are hardened steel and selfaligning. The weighing mechanism is similar, but heavier, in design to Model 4175 listed on SD-WA.87, and the over and under

Testing: The standard tests for an equal arm scale equipped with an over and under head shall apply.

C. S. Phillips,

Assistant Director (W&M),

heads are the mechanical type.

Standards Division.

Director. Standards Division.

Reference A-131.



