

TRADE AND COMMERCE CANADA

## STANDARDS DIVISION

OTTAWA, February 4th, 1959.

APPROVAL LISTING

SINGLE ANIMAL SCALE

## CANADIAN SCALE COMPANY

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform beam scale with or without animal rack, Model 704, manufactured by Canadian Scale Company, 305 Horner Avenue, Toronto 14, Ontario.

Rating of Apparatus: 2500 lbs. capacity.

Application: General farm use and animal weighing.

Conditions: As prescribed in P.C. 6894 and the added requirement that the

scale frame be mounted on a suitable concrete foundation or

secured and mounted on a level concrete floor.

Description: This device is a platform scale equipped with a single bar

notched beam and proportional weights. The minimum beam graduation is 1/2 lb. The fulcrum bearing stands are attached to the fabricated steel frame and the bearings are self-aligning. There are 4-double ball checking units linked at each end of the scale, which have the load bearing blocks fitted to the under section of the ball retaining unit. The levers are equipped with nose irons and are linked to the connecting rod by conical pivots and bearing plates fitted

to shackles. The scale may be used with or without a cattle

rack

Testing The standard tests for a platform scale shall apply.

C.S. Phillips,

Assistant Director (W&M),

Standards Division.

R.W. MacLean, Director, Standards Division.

Reference: A-513

LEVERAGE SYSTEM

SCALE WITH ANDMAL BACK

100 E 704

CANADIAN SCALE COMPANY

General Farm USE and Animal Scale

360