SD-WA. 364

STANDARDS DIVISION

OTTAWA February 18, 1959.

APPROVAL LISTING

SANITARY-BIZERBA - PREPACKAGING SCALE

Under the provisions of the Weights and Measures Act, Chapter 298, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Prepackaging Computing Counter Scale, Model OP-15-P with a capacity of 15 lbs., manufactured by Sanitary Scale Company, Belvidere, Illinois, U.S.A.

Application:

Weighing commodities for prepackaging purposes.

Conditions: As prescribed in P.C. 6894, with the added mandatory requirement that each scale bears a legend to the effect that the scale is "LEGAL FOR PREPACKAGING CNLY AND NOT FOR DIRECT SALES TO CUSTOMERS".

Description: This device is a counter computing scale employing light beams and a transparent chart whose movement is so synchronized with the scale sovement as to allow two systems of condenser and projection lenses with mirrors to be used for projecting images of weight and total price on a screen. The reading line is a definitive color separation reading edge. The price range, ten cents to \$1.59 per pound is set manually by the two levers on the side of the scale. This device utilizes four main load bearing supports and dual opposed pendulums as the resistant. The chart reticule is mounted directly on one of the pendulum levers.

Testing: The standard tests for a prepackaging counter scale shall apply.

C.S. Phillips,

Assistant Director (Wol),

Standards Division.

Reference: A=518

R.W. MacLean,

Director.

Standards Division.

