

Department of consumer and corporate affairs / Ministère de la consommation et des corporations

STANDARDS BRANCH DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA_89

Revision 1

OTTAWA August 7, 1970

Approved: EXACT WEIGHT OVER/UNDER PLATFORM FLOOR SCALES - TYPES 1120, 1140

manufactured by Exact Weight Scale Co. (Canada) Ltd., 1191 Crestlawn Drive, Mississauga, Ontario

Apparatus Listed: Equal Arm Floor Scales with Over and Under Indicator and Tare Beam

Old Model No.	New Type No.	Indicator Aspect
1120V, 1122V	1140	Vertical
1120, 1122	1120	Inclined 30 degrees

Load Rating: Maximum capacity - 150 pounds or metric equivalent.

Application: The weighing of commodities in trade.

Description: These scales are equal arm balance type with the lever of cast iron or aluminum construction. Weight registration is by means of balance weights. The weight platter is enclosed in a housing with access by means of a door at the rear of the scale. The tare beam is not graduated.

The scale base and housings are of aluminum with anodized or enamel The commodity holder is an aluminum platter fitted with a rear guard finish. plate.

Testing: The standard tests for equal arm scales shall apply.

References: SW-85-E5 SL-102-766

Approval is granted under the Weights and Measures Act, Chapter 292, and Note: Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) Chief.


Weights and Measures Division.

Standards Branch.


Director.

Standards Branch.

EXACT WEIGHT OVER/UNDER PLATFORM FLOOR SCALES TYPES 1120, 1140



TYPE 1120



TYPE 1140