Consommate in et Corporation Canada

Consumer Standards Directorate Direction générale des normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-T118

(Issued in advance of preparation of formal Approval Notice)

First Approved: May 1, 1979

Ottawa, April 15, 1980

Company Requesting Approval - Grandin Weighing Systems Ltd.
5895 Bessette Street
Montreal, Quebec H4S IPI

Type of Device -

Compudraft Series 1000 Full Electronic Bulk Grain Weighing System. This system consists of the instrumentation which is used to control the weighing and handling of grain in a terminal transfer situation. The instrument itself may form part of a system including printers, control devices (solenoids, gates, motors, etc.) and digital computer. The inputs to the device are weight signals from load cells and control signals from hopper sensors and gate solenoids.

Conditions of Approval:

- 1. This approval covers the following installations
 - a) 6 units in Calgary (Canadian Gov't Elevator)
 - b) 8 units in Churchill, Manitoba (National Harbours Board)
 - c) 14 units at Elevator #5 at National Harbours Board in Montreal Harbour. These units include CRT's man. by Intelligence Systems Incorporated. They are designated as series 1000C units.
 - d) 2 units at Elevator #3, Tarte Pier, Montreal Harbour.
 - e) 4 units in Quebec City.
 - f) 3 units at Parrish & Heimbacker in Thunder Bay.

File No. G6922-G337

Jey. Buchanan Acting Chief

Weights & Measures Division

Legal Metrology Branch