



Ottawa

24-9-81

NOTICE OF APPROVAL – AVIS D'APPROBATION

Company: W.&T. Avery Limited
Smethwick, Warley
West Midlands
B662LP, England

Société: W.&T. Avery Limited
Smethwick, Warley
West Midlands
B662LP, England

Manufacturer: W.&T. Avery Limited

Fabricant: W.&T. Avery Limited

Type of Device: Portable Platform
Dial Scale.

Appareil: Bascules mobile à tablier
et à cadran.

Model Nos.:

Numéro de modèle:

1) 3205COE - Tare beam is enclosed and has knob-operated tare poise and a tare indicating chart.

1) 3205COE - Fléau de tare encastré comportant un bouton de curseur et un tableau indicateur de tare.

2) 3205COJ - Tare beam is externally mounted and is operated in the conventional manner.

2) 3205COJ - Fléau de tare monté à l'extérieur, à commande ordinaire.

3) 3205CLE - Earlier revision of scale with externally-mounted tare beam.

3) 3205CLE - Bascule décrite dans la révision antérieure comportant un fléau de tare monté à l'extérieur.

Description: These scales utilize two main levers of "A" form and cast construction. One of the main levers has an extension arm which provides a connection to the steelyard rod.

Description: Ces balances utilisent deux leviers principaux coulés en forme de "A" dont un comporte un bras de prolongation le reliant à la tringle d'attaque.

Scale model 3305COE features an optional enclosed rapid tare mechanism. The indicating heads are equipped with Avery type CLA spring resistant mechanism as approved on S.WA-387. The chart reading line diameter is 24 in. on the model 3205COE and COJ and 23 in. on the model 3205COE. Scales may be equipped with single or dual tare beams.

Le modèle 3305COE présente, en option, un mécanisme de tarage rapide. Les cadrans sont équipés d'un mécanisme à ressort CLA de type Avery approuvé en vertu de l'avis S.WA-387. Le diamètre du cadran de lecture des modèles 3205COE et COJ est de 24 po et celui du modèle 3205CLE est de 23 po. Les balances peuvent être munies de fléaux de tare simples ou doubles.

Description: (Cont'd)

The base, load platform and dial column are of cast iron.

The scale may be equipped with a guard frame at the rear of the platform in order to prevent weighed material from contacting the column.

Accessories: Graduated rigid tare devices (model 3205CLE only), back and front indication, load racks and animal pens.

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-A221-23

Description: (Suite)

Le socle, le tablier et la colonne du cadran sont en fonte.

Les balances peuvent être équipées d'une barre montée à l'arrière du tablier empêchant que les marchandises pesées ne touchent à la colonne.

Accessoires: Dispositifs gradués de blocage de la tare (modèle 3205CLE uniquement), affichage à l'avant et à l'arrière, rateliers de charge et enclos pour les animaux.

Conditions d'approbations: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement et de toutes les conditions particulières formulées dans le présent avis.

Référence No: G6922-A221-23

R.G. Knapp
Richard G. Knapp
Director Legal Metrology
Directeur de la Métrology Légale