



Consumer and
Corporate Affairs Canada
Legal Metrology

Consommation
et Corporations Canada
Métrieologie légale

APPROVAL NO. - N° D'APPROBATION
S.WA-T260

APR 20 1983

NOTICE OF TEMPORARY APPROVAL

AVIS D'APPROBATION TEMPORAIRE

Company: Double T Equipment
Manufacturing Ltd.,
P.O. Box 1500,
Airdrie, Alberta
TOM OBO

Société: Double T Equipment
Manufacturing Ltd.,
P.O. Box 1500,
Airdrie, Alberta
TOM OBO

Manufacturer: Double T Equipment
Manufacturing Ltd.,
Airdrie, Alberta

Fabricant: Double T Equipment
Manufacturing Ltd.,
Airdrie, Alberta

Type of Device: Weigh Hopper Blender
scale.

Appareil: Bascule à trémie mélangeuse.

Model Number: DTE6S

No. de modèle: DTE6S

Capacity: 6,500 kg. or avoirdupois
equivalent.

Capacité: 6 500 kg or l'équivalent en
avoirdupois.

Load Cell Capacity: 10000 lb.

Capacité des dynamomètres: 10000 lb

Accessories: Compatible accessories that
meet the requirements and perform in
accordance with the Weights and Measures
Act, Regulations and Specifications may
be used.

Accessoires: Peuvent être utilisés les
accessoires compatibles qui satisfont
aux exigences de la Loi sur les poids et
mesures, du Règlement et des spécifica-
tions y afférents.

Description: This device utilizes a
weigh hopper suspended from 3 electronic
load cells which are connected to a
summing box. The summing box is
interfaced to an electronic digital
indicator.

Description: Il s'agit d'une trémie de
pesage suspendue à 3 dynamomètres élec-
troniques qui sont connectés à une boîte
de sommation. Cette boîte est elle-même
réliée à un indicateur numérique
électronique.

Some models will be equipped with an
agitator in the hopper for blending the
product. The agitator, must be inopera-
tive during the weighing process.

La trémie de certains modèles est munie
d'un agitateur destiné à mélanger le
produit pesé.

Conditions of Approval:

Conditions particulières:

(1) This approval covers 10 devices
only. Each device sold, leased or
otherwise disposed of, must be reported
to the Head, Mass Laboratory, Legal
Metrology Branch, Ottawa, by the
manufacturer or distributor.

(1) La présente approbation ne vise que
10 appareils. Le fabricant ou le
distributeur doit faire part de tout
appareil vendu, loué ou cédé de quelque
autre façon, au Chef du laboratoire des
masses, de la Direction de la Métrologie
légale, à Ottawa.

- 2 -

Conditions of Approval: Continued

(2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

(3) This approval to expire two year from date of issue.


(4) On expiry of this approval only models listed or described having the same nominal capacities and the same nominal physical dimensions and construction as an installation that has been inspected and verified by Weights and Measures, will be considered for full formal approval. After the date of expiry of this Temporary Approval if a model that has not been approved is to be installed, a new application must be submitted, with installation drawings, in order that the Temporary Approval may be reactivated and an inspection arranged.

Conditions d'approbation: Suite

(2) Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

(3) Le présent avis expire deux ans après sa date d'émission.

(4) Dès expiration du présent avis, seuls les modèles dont les capacités et dimensions nominales ainsi que la construction sont identiques à celles d'une installation déjà inspectée et vérifiée par la Division des poids et mesures seront considérés pour approbation officielle entière. Après la date d'expiration de l'approbation provisoire pour qu'un modèle non approuvé soit installé, une nouvelle demande devra être présentée avec les plans de l'installation afin que l'approbation provisoire puisse être de nouveau en vigueur et qu'une date d'inspection puisse être fixée.

Reference No.: G6922-D416No. de référence: G6922-D416


W.R. Virtue

Chief
Legal Metrology Laboratories

Chef
Laboratoires de la Métrologie légale

 20 1983