



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métreologie légale

APPROVAL NO. - N° D'APPROBATION

S.WA-T265

May 17 1983

NOTICE OF TEMPORARY APPROVAL

AVIS D'APPROBATION TEMPORAIRE

Company: Newfoundland Scale Co. Ltd.
P.O. Box 269
Paradise, Newfoundland
AOA 2E0

Société: Newfoundland Scale Co. Ltd.
P.O. Box 269
Paradise, Newfoundland
AOA 2E0

Manufacturer: Easteel Industries
St. Johns, NFLD

Fabricant: Easteel Industries
St. Johns, NFLD

Type of Device: Electronic tank scale.

Appareil: Bascule à cuve électronique.

<u>MODEL NUMBER</u> <u>No. de modèle</u>	<u>CAPACITY</u> <u>Capacité</u>	<u>LOAD CELL TYPE</u> <u>Type de dynamomètre</u>
NST001	4000 LB	De Cisaillement,
NST002	8000 LB	H.B.M Shear Beam
NST003	12 000 LB	load cells/dynomometre
NST004	16 000 LB	
NST005	20 000 LB	
NST010	40 000 LB	
NST020	80 000 LB	
NST035	140 000 LB	
NST050	200 000 LB	
NST075	300 000 LB	
NST100	400 000 LB	
NST200	800 000 LB	
NST300	1 200 000 LB	

*or metric equivalent

*ou l'équivalent métrique

Description: This device is a full electronic tank scale. The tank legs are fitted to four(4) modules which utilize shear beam load cells.

Description: Cet appareil est une balance à cuve entièrement électronique. Les pieds de la cuve sont fixés à quatre (4) modules comportant des dynamomètres de cisaillement.

Weight indication is by means of an approved and compatible electronic digital weight indicator.

L'affichage du poids est assuré par tout indicateur pondéral numérique électronique compatible approuvé.

.../2

Canada

CCA-2067 (4-82)

Conditions of Approval:

(1) This approval covers 10 devices only. Each device sold, leased or otherwise disposed of, must be reported to the Chief, Mass Laboratory, Legal Metrology Branch, in Ottawa, by the Manufacturer or distributor.

(2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

(3) This approval to expire two year from date of issue.

(4) On expiry of this approval only models listed or described having the same nominal capacities and the same nominal physical dimensions and construction as an installation that has been inspected and verified by Weights and Measures, will be considered for full formal approval. After the date of expiry of this Temporary Approval if a model that has not been approved is to be installed, a new application must be submitted, with installation drawings, in order that the Temporary Approval may be reactivated and an inspection arranged.

Reference No.: G6922-N452



W.R. Virtue

Chief
Legal Metrology Laboratories

Conditions d'approbation:

(1) La présente approbation ne couvre que 10 appareils. Le fabricant ou le distributeur doit faire part de tout appareil vendu, loué ou cédé de quelque autre façon, au Chef du laboratoire des masses, de la Direction de la Métrologie légale, à Ottawa.

(2) Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

(3) Le présent avis expire deux ans après sa date d'émission.

(4) Dès expiration du présent avis, seuls les modèles dont les capacités et dimensions nominales ainsi que la construction sont identiques à celles d'une installation déjà inspectée et vérifiée par la Division des poids et mesures seront considérés pour approbation officielle entière. Après la date d'expiration de l'approbation provisoire pour qu'un modèle non approuvé soit installé, une nouvelle demande devra être présentée avec les plans de l'installation afin que l'approbation provisoire puisse être de nouveau en vigueur et qu'une date d'inspection puisse être fixée.

No. de référence: G6922-N452

Chef
Laboratoires de la Métrologie légale

MAY 17 1983