



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métrieologie légale

APPROVAL No. / N° D'APPROBATION

S.WA-T324

DEC 13 1984

NOTICE OF TEMPORARY APPROVAL

AVIS D'APPROBATION TEMPORAIRE

Company: Eastern Scale Mfg. Inc.
30 Carnforth Road,
Toronto, Ontario
M4A 2K9

Manufacturer: Eastern Scale Mfg. Inc.
Toronto, Ontario

Type of Device: Electronic low profile weighing platform that when interfaced to an approved and compatible digital weight indicator becomes a weighing system.

Model Number: ES-312-LP

Capacity: 8000 lbs or metric equivalent.

Platform Size: 12 ft x 3 ft.

Load Cell Capacity: 5000 lbs.

Accessories: Compatible accessories that meet the requirements and perform in accordance with the Weights and Measures Act, Regulations and Specifications may be used.

Description: This device is a fully electronic low profile weighing platform. It utilizes four shear beam load cells mounted on corner stands beneath the platform. Horizontal movement is restricted by 4 bumper bolts mounted on the four corner stands.

A sealable junction box is positioned on the weighbridge.

Société: Eastern Scale Mfg. Inc.
30 Carnforth Road
Toronto, Ontario
M4A 2K9

Fabricant: Eastern Scale Mfg. Inc.
Toronto, Ontario

Appareil: Plate-forme de pesage électronique, surbaissée, qui, une fois reliée à un indicateur pondéral numérique, approuvé et compatible, forme un ensemble de pesage.

N° de modèle: ES-312-LP

Capacité: 8000 lb ou l'équivalent métrique

Dimensions du tablier: 12 pi x 3 pi.

Capacité des cellules de pesage: 5000 lb

Accessoires: Peuvent être utilisés les accessoires compatibles qui satisfont aux exigences de la Loi, du Règlement et des prescriptions ministérielles régissant les poids et mesures.

Description: Il s'agit d'une plate-forme de pesage surbaissée, entièrement électronique. Elle est constituée de quatre cellules de pesage de cisaillement reposant sous la plate-forme, sur des supports placés dans les coins. Le déplacement horizontal de la plate-forme est restreint par 4 boulons de butée fixés aux quatre supports.

Une boîte de jonction scellable est installée sur le tablier.

APPROVAL:

This approval covers 10 devices only. Each device sold, leased or otherwise disposed of, must be reported to the Head, Mass Laboratory, Legal Metrology Branch, in Ottawa, by the manufacturer or distributor.

This approval to expire two years from date of issue.

No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

APPROBATION:

La présente approbation ne vise que 10 appareils. Le fabricant ou le distributeur doit faire part de tout appareil vendu, loué ou cédé de quelque autre façon, au chef du Laboratoire des masses, Direction de la Métrologie légale, à Ottawa.

La présent approbation expire deux ans après sa date d'émission.

Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

Chief
Legal Metrology Laboratories


W.R. Virtue

Chef
Laboratoires de la Métrologie légale

DEC 13 1984