STANDARDS BRANCH

5-WA.391

Ottawa, December 29, 1959

APPROVAL LISTING

SKINNER-SINGLE ANIMAL LIVESTOCK SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform beam scale Model S-38-SA manufact-

ured by Skinner Scales Limited, 51 West Cordova Street,

Vancouver, British Columbia.

Rating of Apparatus: 3500 pounds capacity.

Application: General farm use and animal weighing.

As prescribed in P.C. 6894 and the added

requirement that the scale be mounted on a suitable permanent foundation if located in the open or mounted and anchored on a level concrete floor or concrete slab

within a building.

on: This device is a platform scale with the lever system contained within a channel iron frame. The Description: scale may be equipped with a type registering beam or a full capacity double bar beam with one bar graduated and notched to 3,000 lbs. x 100 lbs. The second bar has a section graduated to 100 lbs. x 1 lb. and one section graduated to 500 lbs. x 2 lbs.

The under deck leverage system consists of two pipe levers with single link and cross bar suspension and two side arms connected to the connecting rod. The scale is also equipped with a shelf lever.

Testing:

The standard tests for a platform scale

shall apply.

C. S. Phillips,

Assistant Director, (W&M),

Standards Branch.

Reference: A-543

& Lunfam R. W. MacLean,

Director,

Standards Branch.

SKINNER - SINGLE ANIMAL SCALE

MODEL S-38-SA

with type registering beam

