

S-WA. 415

STANDARDS BRANCH

OTTAWA, January 26, 1961

APPROVAL LISTING

THORNTON - PORTABLE PLATFORM SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Portable floor type platform scale model 600 LW manufactured by F. J. Thornton and Company Limited, Birmingham, England.

Rating of Apparatus: 2,000 pounds capacity.

Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is a portable beam type platform scale with single bar beam and counterpoise weights. The load platform, base and weigh bridge frame are cast iron. The main lever system consists of two A type levers, one having an extension to the beam rod. The beam is graduated to 100 lbs. x 1/2 lb.

Testing: The standard tests for a platform scale shall apply.

C. S. Phillips, Chief (W&M),

Standards Branch

R. W. MacLean Director,

Standards Branch.

Reference A-589

