



Consumer and  
Corporate Affairs Canada  
Legal Metrology

Consommation  
et Corporations Canada  
Métrieologie légale

APPROVAL No. - N° D'APPROBATION

S.WA-469 Rev. 1

MAY  
MAI

6 1983

## NOTICE OF APPROVAL

Company: Toledo Scale, Division of  
Reliance Electric Limited  
5220 Creekbank Road  
Mississauga, Ontario  
L4W 1X1

Manufacturer: Toledo Scale, Division of  
Reliance Electric Ltd.,  
Mississauga, Ontario

Type of Device: Bench platform dial  
scale.

MODEL NUMBER  
No. de modèle

2020

## AVIS D'APPROBATION

Société: Toledo Scale, Division of  
Reliance Electric Limited  
5220 Creekbank Road  
Mississauga, Ontario  
L4W 1X1

Fabricant: Toledo Scale, Division of  
Reliance Electric Ltd.,  
Mississauga, Ontario

Appareil: Balance de table à plate-  
forme et à cadran.

CAPACITY  
Capacité

50 lb x 1 oz - 20 kg x 50 g  
100 lb x 2 oz - 50 kg x 100 g  
200 lb x 4 oz - 100 kg x 200 g

Description: This device is a bench  
type platform dial scale with a spring  
resistance head mechanism.

The leverage system consists of two  
second class cast iron or fabricated  
steel levers, suspended from the cast  
iron or fabricated steel base. One  
lever is used as a transfer lever.

Conditions of Approval: Approval is  
granted under the Weights and Measures  
Act, S.C. 1970-71-72, chapter 36, and the  
Weights and Measures Regulations  
C.R.C.c., 1605 for use in Canada under  
the general conditions of the said  
Regulations, and under any special  
conditions listed above.

Reference number: G6922-T170-10

W.R. Virtue

Chief  
Legal Metrology Laboratories

Canada

MAY  
MAI

6 1983

APPROVAL No. - N° D'APPROBATION

S.WA-469 Rev. 1

MAY  
MAI

6 1983

Description: Le présent appareil  
consiste en une balance de table à plate-  
forme et à cadran munie d'un mécanisme  
de tête dont la résistance est assurée  
par un ressort.

L'ensemble de leviers consiste en deux  
leviers en fonte ou en acier marchand du  
deuxième genre suspendus au socle en  
fonte ou en acier marchand. L'un des  
leviers sert de levier intermédiaire.

Conditions d'approbation: L'approbation  
est accordée conformément à la Loi sur  
les poids et mesures S.R.C. 1970-71-72,  
chapitre 36, et au Règlement d'appli-  
cation C.R.C.c., 1605. L'emploi est  
autorisé au Canada sous réserve des  
conditions générales dudit Règlement, et  
de toutes les conditions particulières  
formulées dans le présent avis.

No. de référence: G6922-T170-10

Chef  
Laboratoires de la Métrologie légale

CCA-873 (4-82)