Consommation et corporations

Legal Metrology

Mètrologie Lègale

S.WA-481

Approval No. d'approbation Revision 1

First approved Dec. 2, 1963

Ostawa August 12, 1980

NOTICE OF APPROVAL - AVIS D'APPROBATION

DETECTO - PORTABLE PLATFORM FLOOR SCALES

MANUFACTURER: D

Detecto Scales Company

103-00 Foster Ave. Brooklyn, New York

TYPE OF DEVICE: Portable platform floor scales with mechanical dial and/or beam indication, or with electronic digital readout in conjunction with a load cell.

	Model Series		Max. Capacity	
Beam	Mech. Dial	Digital	Pounds	Kilograms
854			1000	500
	2300	8230	2230	1000
~	· ·	8850	1000	400

APPLICATION: The weighing of commodities in trade.

DESCRIPTION: Model 854 scale is equipped with two A-shaped load levers one of which has an extension which connects to the steelyard. Indication is by means of single or double beams in conjunction with counterpoise weights.

Series 2300 scales have two T-shaped load levers connected to a transfer lever and utilize a double pendulum head mechanism and are equipped with tare and capacity beams.

Digital readout series 8230 and 8850 have the same lever systems as series 2300 and 854 respectively, with weight sensing by means of a tension load cell installed at the steelyard.

FILE: G6922-D165-18

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36 and the Weights and Measures Regulations (C. 1605) for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.

John Armstrong

Chief, Weights and Measures Legal Metrology Branch

S.WA-481 Revision 1 August 12, 1980

