

CANADA

## DEPARTMENT OF TRADE AND COMMERCE

WEIGHTS AND MEASURES DIVISION

OTTAWA, June 5, 1964.

## APPROVAL LISTING

## TOLEDO VERILUX PULL INDICATION

Under the provisions of the Weights and Measures Act, Chapter 222, R.S.C. 1952 and Regulations thereunder (R.R. 609), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

**Apparatus Listed:** Toledo Verilux "Pull" Indication is manufactured by the Toledo Scale Company of Canada Limited, 2462 Howard Avenue, Windsor, Ontario. This device is optional on all approved Toledo Scale Models in the 2000 series.

**Application:** Weighing commodities in general trade and for industrial use.

**Conditions:** As prescribed in P. C. 609.

**Description:** This device provides a projection type indication with the weight shown on a viewing screen. The Verilux indication utilizes a chart wheel instead of the conventional type pointer indicator. The chart wheel of thin metal is held in position on the pinion shaft by means of a collet and a lock nut.

The chart is a strip of plastic one inch wide which is fastened to the chart wheel. By means of a light source, a lens and a series of four reflecting mirrors, the chart reading is projected on a glass screen.

**Testing:** The standard tests for dial scales shall apply.

*C. S. Phillips*  
C. S. Phillips,  
Chief, Weights & Measures Division,


Standards Branch.

*R. W. MacLachlan*  
R. W. MacLachlan,  
Director,  
Standards Branch.

Ref: SI-102-732

(

)



TOLEDO VERILUX "711" INDICATION