

# DEPARTMENT OF TRADE AND COMMERCE MINISTÈRE DU COMMERCE

S-WA.581

STANDARDS BRANCH

... November 15. 1965

### APPROVAL LISTING

## BERGLUND SWIFTWEIGHT OVER AND UNDER COUNTER SCALE

#### Model E.S.A.-10

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C.1952, and Regulations thereunder (P.C.6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus listed:

Equal arm counter scale with over and under indication head,

Model E.S.A.-10.

Manufactured by Industrial Scale Company, 5771 N.E. Union

Avenue, Portland 11, Oregon.

Rating of Apparatus:

Capacity 10 pounds or metric equivalent.

Application:

Weighing commodities in general trade.

Conditions:

As prescribed in P.C.6894. The over and under chart may be graduated, but weight values must not be marked on the

chart.

Description:

This device is an equal arm counter scale with over and under end tower head indication. The side beam is notched and graduated to 16 ounces by  $\frac{1}{4}$  ounce, or 500 grams by 10 grams. The load receiving element may be either a platter or a scoop. The single lever is an aluminum alloy casting. The pivots are pear shaped hardened steel and the bearings may be either V notched agate blocks or V notched hardened steel. The stabilizing mechanism is of the double bar and pin type. The base, column and over and under head

casing are polished aluminum castings.

Testing:

The standard tests for an equal arm counter scale shall

apply.

J. L. Armstrong,

Chief, Weights and Measures Division,

Standards Branch.

R.W. Maolean,

Director,

Standards Branch.

Ref: SL-102-579

# BERGLUND SWIFTWEICHT OVER AND UNDER COUNTER SCALE MODEL E.S.A.-10



