

Department of consumer and corporate affairs / Ministère de la consommation et des corporations

## STANDARDS BRANCH - DIRECTION DES NORMES

## NOTICE OF APPROVAL

S.WA-773

OTTAWA December 5, 1969.

Approved: AVERY FAN TYPE COUNTER SCALE - MODEL 1218CJB

manufactured by W. & T. Avery Ltd., Birmingham 40, England. Canadian distributor: Robert Morse Corp. Ltd., Dominion Industrial Building, Ville St. Fierre, Quebec.

Apparatus Listed: Fan Type Counter Scale, Model 1218CJB. Front and Rear Indication. Charts may be graduated for weight only or for weight and price computing. A tare beam is optional.

Load Rating: Maximum capacities are - Avoirdupois - 4 lb., Metric - 2 kg.

Application: The weighing of commodities in trade.

Description: This scale utilizes a single load lever of cast iron construction with parallel link stabilizer. The stabilizer link is mounted above the load lever. Fulcrum points are fitted with ball bearing assemblies.

The indicator head is equipped with a cam and single pendulum mechanism. Additional resistance is provided by means of two helical tension springs, one of which is so connected as to resist cam rotation, the other being connected between the forward end of the load spider and the scale base.

The scale base and housings are of cast construction with enamel finish.

The scale may be equipped with a pan or scoop of stainless steel or plastic.

Testing: The standard tests for counter scales shall apply.

References: SW-85-A3 SI-102-355

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

Chief, Weights & Measures Division, Standards Branch.

Director,

Standards Branch.





