

Department of consumer and corporate affairs / Ministère de la consommation et des corporations

STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-799

OTTAWA July 30, 1970

Approved: TOLEDO POSTAL SCALE - MODEL 3121P

manufactured by Toledo Scale Co. of Canada Ltd., 2462 Howard Avenue, Windsor, Ontario

Apparatus Listed: Fan Type Postal Scale with Weight and Price Indication, Model 3121P.

Load Rating: Maximum capacity - 10 pounds.

Application: Mail weighing and the determination of postal charges.

<u>Description:</u> This scale utilizes a single load lever with ball-type bearings. Load resistance is by means of a single cam and pendulum assembly which carries the indicator pointer.

The indicator makes two complete passes over the chart in order to register capacity load. When half load is reached a pair of externally-mounted counterweights are manually rotated to act on the load lever, resulting in the indicator being returned to chart zero. The chart has two complete sets of weight graduations, from zero to one-half load and from half load to capacity load, with identification by means of contrasting colours.

Except for the double-travel indicator feature this scale is the same as Toledo Model 3111 (S.WA-640).

Testing: The standard tests for fan type counter scales shall apply.

References: SW-85-T6 SL-102-106

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for)Chief,


Weights and Measures Division,

Standards Branch.

R.W. MacLean,

Director.

Standards Branch.



. . .