

Department of consumer and corporate affairs/Ministère de la consommation et des corporations

SYANDARDS BRANCH - DIRECTION DES NORMES



NOTICE OF APPROVAL-AVIS D'APPROBATION

NO.	***************************************	
DATE	S.WA-820	
	August 23,	1971

OHAUS OVER-UNDER BENCH SCALES - SERIES 8800

Manufacturer:

Ohaus Scale Corporation Florham Park, New Jersey,

U.S.A.

Apparatus: Over-Under Scales, with graduated beam and chart incication, model series 8800. The indicating chart may be graduated and printed with either avoirdupois or metric weight values. A tare beam is optional.

Rating:

Maximum capacity - 10 pounds. 5 kilograms.

Application:

The weighing of commodities in trade.

Description: This scale is equipped with an equal arm cast alloy lever connected to the over-under indicator by means of a pullrod. Platform load is resisted by means of balancing weights, in conjunction with a graduated beam and poise.

Scale adjustment is effected by adjusting the lever applied by a coil spring which is connected to the scale lever. The scale is equipped with a magnetic dampening system.

Scale housings are of cast alloy construction. Load and weight platters are of stainless steel.

Testing:

The standard tests for over-under scales shall apply.

References:

SW-85-03 SL-102-371

Conditions Of Approval:

Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P. C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

for) Chief,

Weights and Measures Division

Standards Branch

H. W. MacLeart

Director,

Standards Branch

OHAUS OVER-UNDER BENCH SCALES - SERIES 8800

