

Department of consumer and corporate affairs / Ministère de la consommation et des corporations



### STANDARDS BRANCH - DIRECTION DES NORMES

# NOTICE OF APPROVAL-AVIS D'APPROBATION

S.WA - 827

DATE

January 8, 1972

#### HOWE RICHARDSON PLATFORM SCALES - DIAL INDICATION

Manufacturer:

Robert Morse Corporation Limited

Dominion Industrial Building

Ville St. Pierre, Quebec.

Apparatus:

Platform Dial Scales

Model	Platform Size	Max. Capacity	
Number	Inches	Pounds	Kilograms
6401 6402	19 x 28 28 x 28	1000 2000	500 1000

Scale indication may be in pounds or in

metric equivalent.

Scales may be furnished with either a short or a long column, for bench or floor use respectively.

Scales up to 500 lb. capacity may be

furnished with counting attachments having ratios of 50:1 and 10:1 or 49:1 and 9:1.

Scales of capacity above 500 lb. may be furnished with counting attachments having ratios of 100:1 and 10:1 or 99:1 and 9:1.

Rating:

Maximum capacities are as listed above.

Application:

The weighing of commodities in trade.


Description: These scales are equipped with two load levers fabricated from steel bars and pipe and assembled by welding. The dial head mechanism is double pendulum type with tape drive to the indicator, commonly known as Howe-Richardson Magna-Weigh type.

The scale platforms and housings are constructed of sheet steel with enamel finish. Beams are of stainless steel with die cast poises. Floor scales are equipped with wheels.

S.WA-827

## HOWE RICHARDSON PLATFORM SCALES - DIAL INDICATION





6402

Testing:

The standard tests for platform scales

shall apply.

References:

SW - 85 - H11

SL -102 - 88

#### Conditions of Approval:

Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P. C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) / Chief. Weights and Measures Division Standards Branch

Director,

Standards Branch