Consommation et corporations

Standards

Normes

NOTICE OF APPROVAL AVIS D'APPROBATION

S.WA-848

Ottawa October 2, 1975

TOLEDO ONE-SPOT READING DIAL

Manufacturer:

Toledo Scale Division,

Reliance Electric Limited,

2462 Howard Ave., Windsor, Ontario

Apparatus:

Toledo One-Spot Reading Dial, Model 701

Rating:

Capacity - 50 lb. to 100,000 lb.

Application: For use as an alternative for the normal dial and pointer on Toledo 20" mechanical dial heads, with or without unit weights.

Description: The Toledo One-Spot system utilizes a rotating dial in conjunction with a fixed indicating line. Weight indication is shown through a magnifying lens and internal lighting is provided, which permits reading the scale from greater than normal distance.

On scales equipped with unit weights, the value of active unit weights is automatically displayed in a window which is adjacent to the dial weight indicating line.

Testing: shall apply.

The standard tests for dial indicating scales

Reference:

GL1153-57/T170-392

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

MacLean Director General

Standards Directorate

TOLEDO ONE-SPOT READING DIAL

