RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA-924 (Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Company Requesting Approval -Western Scale Co. Ltd. Port Coquitlan, B.C.

Type of Device

Western model DS300-T2 Electronic Digital

Indicator for Western Lever' Type scales

Model No.

Rating or Capacity

Remarks ·

DS-300-T2

Western Digital Indicator manufactured for Western Scale Co. Ltd. by Doric Scientific Corp. San Diego, Cal. U.S.A.

Accessories:

Remote Indicator and Weight Printer must meet all the requirement and perform in accordance with the Weights and Measures Act

and Regulations.

Application:

The Western Model DS300-T2 Digital Indicator with one or moreelectronic load cells may be used with approved Western lever type load receivers which are compatable.

Date of Application for Approval -

File No. 1153-57/W172-644

October .22, .1975 Date of Filing

Signature

File Record of Approval Action with Approval Listing notices when Moval Listing number is assigned in advance of preparation of notice; file a of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listin notice with printed notice when available.

Consumer and Corporate Affairs

Consommation et corporations

Standards

Normes

NOTICE OF APPROVAL **AVIS D'APPROBATION**

S.WA-925

July 19, 1976

OHAUS GRAIN TEST SCALE - MODEL 14M

Manufacturer:

Ohaus Scale Corporation

Florham Park, New Jersey, U.S.A.

Apparatus:

Grain Test Scale - Model 14M

Rating:

Weighing capacity - 550 grams

Application:

Direct weighing of grain test samples Determining moisture content in grain

Determining density of grain

Description:

This is an unequal arm, single lever scale equipped with two graduated beams and a plastic commodity

scoop.

This scale is identical with Ohaus Model 14A (S.WA-541), except that Model 14M is graduated in metric units and has a commodity scoop which differs in form and material from that used on the earlier model.

Testing:

The standard tests for grain test scales

shall apply.

Reference:

G6900-069-3

Conditions of Approval: Approval is granted under the Weights and Measures Act. S.C. 1970-71-72, Chapter 36, of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Director General

Consumer Standards Directorate

OHAUS GRAIN TEST SCALE - MODEL 14M

