RECORD APPROVAL ACTION

Approval Listing Number - SWA-1021

(Issued in advance of preparation of draft Approval Listing only in case of

Interim Approval).

Western Scale Company Limited

1670 Kingsway Avenue Company Requesting Approval -

Port Coquitlam, British Columbia V3C 3Y9

Type of Device - Digital Electronic Weight Indicator Models DF100 and DF110

Rating or Capacity Remarks Model No. DF100 Approved Capacity of the weight DF110 receiving element.

The Western Digital Indicator models DF100 and DF110 with Application:

one or more approved type load cells may be used with approved type load receiver which are compatible.

Auto zero, compatible printer, digital tare, motion Options:

detector (for printer inhibit).

Note: Models DF100 and DF110 to be equipped with a test button located on the front of the instrument when depressed will activate the L.E.D. to show eights

on all segments.

Model DF110 to be supplied with power interrupt, the ΙI weight display will blank out and it will be necessary to re-zero the scale when power is re-applied before the unit will operate.

Model DF110 to be equipped with a manually-operated III slotted shaft zero adjust control on the front face of the indicator.

IV Model DF110 to be programmed to ensure that auto zero tracking (AzT) operates at a maximum tracking value of one minimum increment of registration.

Date of Application for Approval -

Central File No. - G6900-W172

Schedule I - 275 and 276

Lab. File No. -

September 6, 1977 Date of Filing.

File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices, replace draft of Approval Listing notice with printed notice when available.