

Consommation et corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA 1099

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Ottawa Sept. 18, 1978

Company Requesting Approval - Canadian Laboratory Supplies 80 Jutland Rd., Toronto, Ont. H8Z2H4

Type of Device

Sartorius- Mechanical Top loading balances.

Model 1103 Capacity 2000 g x 0.1 g

Model 2354 Capacity 1000 g x 0.01 g

The design features of the above two scales are similar, but their capacities are different.

File No. G6922-S 803

∮ohn Armstrong

Chief

Weights and Measures Division Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices, replace draft of Approval Listing notice with printed notice when available.