Consommation et corporations

Standards

Normes

CORRECTED COPY - Nov. 7/78 (see below)

## RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-1100

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

September 19, 1978

Company Requesting Approval - Metro Equipment Corp., - Box 60037, Sunnyvale, California 94088

Type of Device 18" x 18" electronic platform Model AF182(or SAF 182) with single load cell model CBU 1-25-A25.

Capacity 200 lb or 100 kg

Resolution 1 part in 5000 at 100 lb. 400 at 200 lb. should read 4000 l part in 400 at 200 lb

platform - stainless steel.

File No. G6922-Y28

John Armstrong

Chief

Weights and Measures Division

Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.