Consommation et corporations

Standards

Normes

## RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA 1130

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Ottawa, April 10, 1979

Company Requesting Approval -30 Baywood Rd., Unit 14
Rexdale, Ontario, M9V 3Z2

- 1. Manufactured by: City Wide Scale Co. Ltd., Rexdale, Ontario.
- 2. Type of Scale: Digital Electronic Weight meter, portable and with external load cell inputs with net/gross switch
- 3. Use of Scale: Industrial
- 4. Capacity & Readout increments: Maximum 8000 counts displayed with an internal count of 32,000 counts, with following ranges of capacity:
  - a) 8000 x 1.0 lb
  - b)  $800 \times 0.1 1b$
  - c)  $80 \times .01 \text{ lb}$
- 5. Model Number: 5500
- 6. Device options: None
- 7. Device operational features: The device has a 4 digit LED display with selectable decimal point. It has zero and motion indicators. It can accept up to 6 load cell inputs and also includes lb/kg switch on the front panel.

File No. G6922-C793-33

901.

ohn Armstrong

Chief

Weights and Measures Division Legal Metrology Branch

erager

المنية

....