Consommation et Corporations Canada Metrologie legale

RECORD OF APPROVAL ACTION

Approval Listing Number | s.wa-1130 Revision 1

(Issued in advance of preparation of formal Approval Notice)

Ottawa, July 3, 1980
First Approved, April 10, 1979

Company Requesting Approval: City Wide Scale Co. Ltd.
30 Baywood Rd., Unit 14
Rexdale, Ontario M9V 3Z2

- 1. Manufactured by: City Wide Scale Co. Ltd., Rexdale, Ontario.
- 2. Type of Scale: Digital Electronic Weight meter, portable and with external load cell inputs with net/gross switch.
- Use of Scale: Industrial.
- 4. Capacity & Readout Increments: Maximum 10,000 counts displayed with an internal count of 32,000 counts, with following ranges of capacity:
 - 100,000 x 10 1b a) b) $10,000 \times 2 1b$ 5,000 x 1 1b c) d) $2,000 \times 0.5 1b$ $1,000 \times 0.2 lb$ e) 500 🗴 f) 0.1 lb 200 x 0.05 lb g) 100 x 0.02 1ь h) 50 x 0.01 lb i) j) Under 20 x 0.005 lb.
- 5. Model Number: 5500.
- 6. Device Option: None.
- 7. Device Operational Features: The device has a 4 digit LED display with selectable decimal point. It has zero and motion indicators (plus 2 dummy zeros). It can accept up to 6 load cell inputs and also includes lb/kg switch on the front panel.

File No.9 G6922-C793-33

John Buchanan Acting Chief

Weights and Measures Legal Metrology Branch

. . . .