Consommation et Corporation Canada

Ottawa/Hull KlA OC9

Consumer Standards Directorate Direction générale des normes

## RECORD OF APPROVAL ACTION

Approval listing Number - S.WA-1136

(Issued in advance of preparation of draft Approval Listing only in case of interim Approval).

Ottawa, June 4, 1979.

Company Requesting Approval - Sanitary Scale Co.
910E Lincoln Ave.
Belvidere, Illinois 61008, U.S.A.

.

1. Manufactured by: Sanitary Scale Co., Illinois

2. Type of Scale: Prepack scale system with weight scale module and label printer

3. Use of Scale: Retail trade - Prepacked weighing only.

4. Capacity & Readout increment 25 x .01 lbs

5. Model No.: TR2

6. Device options: Totaliser for item count, total money since last clear & total weight (lb or kg) since last clear can be obtained in rapid sequence either by key switch or push button.

7. Device operational features: The device is/micro-processor based weighing system with an Attached label printer.

The Weight sensor is electro-optical with a 5 digit LED display. Both auto tare and digital tare functions are included as regular features of the device.

John Armstrong

Chief

Weights & Measures Division Legal Metrology Branch

File No. G6922-S526-23

		ا مورود
		-
•	 ·· 00000	