



Consumer and
Corporate Affairs
Canada

Legal Metrology

Consommation
et Corporations
Canada

Métrie Légale

S.WA-1150 Add. 1

Approval No. d'approbation

Ottawa

JUL 22 1982

NOTICE OF APPROVAL – AVIS D'APPROBATION

This addendum is to be read in
conjunction with S.WA-1150 Rev. 1

Le présent supplément doit être lu de
concert avec l'avis S.WA-1150,
révision 1.

Company: Toledo Scale
Division of Reliance Electric
5220 Creekbank Road
Mississauga, Ontario
L4W 1X1

Société: Toledo Scale
Division of Reliance Electric
5220 Creekbank Road
Mississauga, Ontario
L4W 1X1

Type of Device: Electronic scales that
may be used as computing scales or as a
checkstand scale.

Appareil: Balances électroniques
pouvant servir de balances calculatrices
ou de balances de comptoir de vente.

<u>MODEL NUMBER</u> <u>No. de modèle</u>	<u>CAPACITY</u> <u>Capacité</u> <u>(ATT modeTs/Tout modèles</u>	<u>LOAD CELL</u> <u>Dynamomètre</u>
8404	30 lb x .01 lb.	Toledo Part No.
8406	15 kg x .005 kg	A10518800A/référence
8408	15 lb x .005 lb. 6 kg x .002 kg.	Toledo A10518800A

The following is a list of ECRs that are
approved for use with these scales:

Voici la liste des caisses enregistreuses
électroniques pouvant être utilisées avec
les présentes balances:

National Semiconductor	T2000
National Semiconductor	MT2000
National Semiconductor	T2500
National Semiconductor	T2600
National Semiconductor	T3000
National Semiconductor	MT3000
Sweda	425
IBM	série 3663 series

Condition of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-T170-23

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: G6922-T170-23

Acting Chief
Legal Metrology Laboratories


W.R. Virtue

Chef Intérimaire
Laboratoires de la Métrologie légale