



Consumer and
Corporate Affairs
Canada

Legal Metrology

Consommation
et Corporations
Canada

Métrieologie Légale

S.WA-3068 Rev. 1

Approval No. d'approbation

Ottawa

SEP 3 1982

NOTICE OF APPROVAL – AVIS D'APPROBATION

THIS NOTICE OF APPROVAL SUPERSEDES RECORD
OF APPROVAL ACTION S.WA-T156 DATED
October 2, 1980 AND S.WA-3068 DATED
March 1, 1982.

Company: Toledo Scales
Division of Reliance Electric
Limited
5220 Creekbank Road
Mississauga, Ontario

Type of Device: Self-contained
electronic floor scales, with weight
sensing by means of cantilever type load
cells supporting the weighbridge, and
readout by means of an approved digital
indicator.

Model Number: 2295/xxxx. The latter
portion of the model number serves to
identify the indicator model.

Platform Size: 4'x6'; 6'x8'; 7'x9';
5'x10'; 6'x10'; 7'x10'; 8'x10'; 7'x11'.

Maximum Rating: 40,000 lb or 20,000 kg.

Condition of Approval: Approval is
granted under the Weights and Measures
Act, S.C. 1970-71-72, chapter 36, and the
Weights and Measures Regulations
C.R.C.c., 1605 for use in Canada under
the general conditions of the said
Regulations and under any special
conditions listed above.

Reference No.: G6922-T170-10

L'AVIS D'APPROBATION REMPLACE L'AVIS
S.WA-T156 EN DATE DU 2 octobre 1980,
S.WA-3068 EN DATE DU 1 mars 1982.

Société: Toledo Scales
Division of Reliance Electric
Limited
5220 Creekbank Road
Mississauga, Ontario

Appareil: Bascule de sol électronique
autonome dans laquelle à la fois éléments
déetecteurs de charge et éléments de
soutien de châssis récepteur.
L'affichage est réalisé au moyen d'un
indicateur numérique approuvé.

Numéro de modèle: 2295/xxxx. La
dernière tranche du numéro de modèle
désigne le modèle de l'indicateur.

Dimensions du tablier: 4'x6'; 6'x8';
7'x9'; 5'x10'; 6'x10'; 7'x10'; 8'x10';
7'x11'.

Portée maximale: 40 000 lb ou 20 000 kg.

Condition d'approbation: L'approbation
est accordée conformément à la Loi sur
les poids et mesures S.R.C. 1970-71-72,
chapitre 36, et au règlement d'application
C.R.C.c., 1605. L'emploi est autorisé
au Canada sous réserve des conditions
générales dudit règlement, et de toutes
les conditions particulières formulées
dans le présent avis.

No. de référence: G6922-T170-10

D. McEvoy

...ing Chief
Legal Metrology Laboratories

Chef intérimaire
Laboratoires de la Métrologie légale