



Consumer and
Corporate Affairs
Canada

Legal Metrology

Consommation
et Corporations
Canada

Métrie Légale

S.WA-3111

Approval No. d'approbation

Ottawa

DEC 6 1982

NOTICE OF APPROVAL – AVIS D'APPROBATION

Company: Metro Equipment Corporation
Box 60037
Sunnyvale, California
USA 94088

Manufacturer: Yamato Scale Co. Ltd.
5-22 Chaemba - Cho
Akashi Hyogo - Pref.
Japan

Type of Device: Digital electronic bench scales, with a single weight only display capability.

Société: Metro Equipment Corporation
Box 60037
Sunnyvale, California
USA 94088

Fabricant: Yamota Scale Co. Ltd.,
5-22 Chaemba - Cho
Akashi Hyogo - Pref.
Japan

Appareil: Balance électronique, de table, à affichage numérique du poids, uniquement.

MODEL NUMBER
No. de modèle

ACS 1611

CAPACITY
Capacité

6 x 0.002 lb / 3 x 0.001 kg
15 x 0.005 lb / 6 x 0.001 kg
30 x 0.01 lb / 15 x 0.005 kg
60 x 0.02 lb / 30 x 0.01 kg
125 x 0.05 lb / 60 x 0.02 kg

Accessories: Compatible accessories that meet the requirements and perform in accordance with the Weights and Measures Act, Regulations and Specifications may be used.

Description: These electronic load cell bench scales are available with optional dual (front and back) weight display configurations. There are LED annunciators for Center of Zero and incorrect Tare.

The following features may be found on the device:

- 1) 10 numeric buttons designated 0 through 9, used for entering tare.
- 2) "TR" button used for achieving a Tare.


Accessoires: Peuvent être utilisés les accessoires compatibles qui satisfont aux exigences de la Loi, du règlement et des spécifications ministérielles régissant les poids et mesures.

Description: Ces balances munies de dynamomètres électroniques sont disponibles avec (en option) un dispositif d'affichage du poids à deux écrans (avant et arrière). Des voyants à DEL s'allument pour indiquer le centre du zéro et une tare incorrecte.

L'appareil réunit les éléments suivants:

- 1) 10 touches numérotées de 0 à 9 qui permettent d'introduire la tare;
- 2) la touche "TR" utilisée pour obtenir une tare;

- 2 -



- 3 -

Description: Continued

3) "C" button used for clearing the scale.

The scale is fitted with a levelling bullseye and four adjustable locking feet.

Provisions for Sealing: The span and coarse zero adjustments can be sealed with a tamper proof paper seal or with a wire through drilled head screws.

Special Conditions: The sealing means is exempt from providing ready access to components or adjustments per SGM3/10.

Condition of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-H186-33

Chief
Legal Metrology Laboratories

Description: Suite

3) la touche "C" qui sert à remettre la balance à zéro.

La balance est équipée d'un indicateur de niveau et de quatre pieds de verrouillage réglables.

Condition de plombage: Les dispositifs de réglage de la portée et du zéro approximatif peuvent être plombés à l'aide d'un scellé en papier inviolable ou au moyen d'un fil métallique traversant des vis à tête perforée.

Condition particulières: Le plombage n'est pas tenu de permettre l'accès facile aux composants et dispositifs de réglage conformément à la prescription ministérielle SGM3/10.

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement, et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: G6922-H186-33

W.R. Virtue

Chef
Laboratoires de la Métrologie légale

DEC 6 1982