



Consumer and
Corporate Affairs Canada
Legal Metrology

Consommation
et Corporations Canada
Métrieologie légale

APPROVAL No. - N° D'APPROBATION
S.WA-3242

APR 22 1985

NOTICE OF APPROVAL

Issued by statutory authority of
the Minister of Consumer and Corporate
Affairs under application by:

Tokyo Electric Canada Limited
9 Meteor Drive
Rexdale, Ontario
M9W 1A3

for the following devices:

**DEVICE TYPE /
TYPE D'APPAREIL:**

Electronic Pre-Pack Scale / Balance de
pré-emballage électronique.

**MODEL DESIGNATIONS /
DESIGNATIONS DES MODELES:**

SL32-25R
SL31-50R

Accordée en vertu du pouvoir stat-
utaire du Ministre de Consommation et
Corporations à la demande de:

pour les appareils suivants:

**MANUFACTURER /
FABRICANT:**

Tokyo Electric Co. Limited

**RATING-CAPACITY-RANCE(S) /
CLASSEMENT-CAPACITE-ETENDUE(S):**

25 kg x .005 kg
or/ou
50 lb x .01 lb.

NOTE: This approval applies only to
devices, the design, composition, con-
struction and performance of which are,
in every material respect, identical to
that described in the information
submitted and are typified by the
sample(s) submitted by the applicant
for evaluation for approval in accor-
dance with sections 14 and 15 of the
Weights and Measures Regulations. The
following is a summary of salient
features only.

REMARQUE: La présente approbation ne
vise que les appareils dont la concep-
tion, la composition, la construction
et le rendement sont identiques, en
tout point, à ceux qui sont décrits
dans la documentation reçue et pour
lesquels des échantillons repré-
sentatifs ont été fournis par le requérant
aux fins de d'évaluation, conformément
aux articles 14 et 15 du Règlement sur
les poids et mesures. Ce qui suit est
une brève description de leurs prin-
ciples caractéristiques.

SUMMARY DESCRIPTION:

The approved device is an electronic pre-pack computing scale. It is a modular format and consists of:

(1) a weight platform that consists of a base frame fitted with 4 levelling feet and a bullseye level, a cantilever load cell is mounted on the frame and it supports a load spider fitted with a stainless steel platter measuring 400mm x 260mm.

(2) an operators digital weight display that can be wall or table mounted and is fitted with the following controls and displays:

(a) Displays for Weight (5 digits), Unit Price (4 digits) and Total Price (5 digits) are of the 7 segment type green in colour and 12.5mm high.

(b) LED annunciators for NET, Pre-Pack, Center Zero, Price/kg, Price/100g, Price/lb, Price/ $\frac{1}{2}$ lb and Price/ $\frac{1}{4}$ lb.

(c) a recessed switch for zero

(d) a bank of keys numbered from 0 to 9 for input of unit price or keyboard tare as well as switches for Clear, P/TR, 100g, Print, S for pre-pack and one marked ETR.PLU.CLR that is used to enter, call up or clear a PLU when connected to a printer with these facilities.

DESCRIPTION SOMMAIRE:

Il s'agit d'une balance de pré-emballage électronique de type modulaire qui comporte les éléments suivants:

(1) une plate-forme de pesage constituée d'un bâti muni de 4 pieds réglables et d'une bulle de mise au niveau et une cellule de pesage de flexion montée sur le bâti et sur laquelle repose une araignée de charge surmontée d'un plateau en acier inoxydable mesurant 400 mm x 260 mm;

(2) un dispositif d'affichage pondéral numérique à l'intention de l'opérateur pouvant être installé au mur ou sur un pupitre qui présente les éléments suivants:

(a) des dispositifs d'affichage verts à 7 segments dont les chiffres mesurent 12.5mm de hauteur servant à indiquer le poids (5 chiffres), le prix unitaire (4 chiffres) et le prix total (5 chiffres);

(b) de DEL associées au poids net, au mode de pré-emballage, au zéro central, au prix/kg, au prix/100g, au prix/lb, au prix/ $\frac{1}{2}$ lb et au prix/ $\frac{1}{4}$ lb;

(c) un commutateur en retrait pour la mise à zéro;

(d) une série de touches numérotées de 0 à 9 servant à introduire le prix unitaire ou la tare de clavier ainsi que différents commutateurs notamment clear (effacement), P/TR (prix/tare), 100g, Print (impression), S (pré-emballage) et ETR.PLU.CLR (servant à introduire, à rappeler ou à effacer un code PLU lorsque l'appareil est relié à une imprimante pouvant effectuer ces fonctions.)

APPROVAL:


The design, composition, construction and performance of the device type(s) identified herein having been evaluated in accordance with regulations and specifications relating thereto, established under the Weights and Measures Act, approval is hereby granted pursuant to subsection 3(1) of the said Act.

The marking, installation, use and manner of use in trade of devices are subject to inspection in accordance with regulations and specifications relating thereto, established under the Weights and Measures Act, and certification of conformity is required in addition to this approval. Inquiries regarding inspection and certification of conformity should be addressed to the local inspection office of Consumer and Corporate Affairs Canada. Requirements relating to marking are set forth in sections 18 to 26 of the Weights and Measures Regulations. Requirements relating to installation, use and manner of use are set forth in Part V and in specifications established pursuant to section 27 of the said Regulations.

APPROBATION:

La conception, la composition, la construction et le rendement des types d'appareils identifiés ci-dessus ayant fait l'objet d'une évaluation aux fins d'approbation conformément au Règlement et aux prescriptions établis en vertu de la Loi sur les poids et mesures, une approbation est accordée par les présentes en application du paragraphe 3(1) de ladite loi.

Le marquage, l'installation, l'utilisation et le mode d'emploi des appareils sont soumis à l'inspection conformément au Règlement et aux prescriptions établis en vertu de la Loi sur les poids et mesures, et doivent être certifiés conformes en sus d'être approuvés par les présentes. Toute demande de renseignements sur l'inspection et la certification de conformité doit être adressée au bureau d'inspection local de Consommation et Corporations Canada. Les exigences de marquage sont définies dans les articles allant de 18 à 26 du Règlement sur les poids et mesures. Les exigences relatives à l'installation, à l'utilisation et au mode d'emploi sont définies dans la partie V et dans les prescriptions établies en vertu de l'article 27 dudit règlement.



W.R. Virtue

Chief
Legal Metrology Laboratories

Chef
Laboratoires de la Métrologie légale

APR 22 1985

File Number / Numéro de dossier:

O6922-T280

